## EXHIBIT G-2 THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

## ARIZONA DEPARTMENT OF PUBLIC SAFETY

## STANDARD QUALITY ASSURANCE PROCEDURES INTOXILYZER MODEL 8000

## STANDARD CALIBRATION CHECK PROCEDURE

DATE (2/6/201) DATE (2/6/201)
DATE [4/6/201] TIME 1853
INTOXILYZER SERIAL # 80-00/858 LOCATION 190 N GTCHETELD RO
LOCATION
Pour a standard alcohol concentration solution
DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: 190 N LITCHFIELD RD
Serial Number: 80-001858
Core Version: 8105.48
12/16/2011

Standard Lot#: 01011100A1
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test

Test

G/210L

Time

Test	g/210L	Time A	)
Air Blank	0.000	18:54:11	
0.100 Cal Che	0.098	18:54:32	
Air Blank	0.000	18:55:02	

Preventative Maintenance Performed -- Timer Reset